| * Frequency is 2.0 – 11.5 MHz xxx xxx xxx | * Minimum Frequency is 1.0–3.0 MHz and Maximum Frequency 11.5 MHz XXX XXX XXX | |
|---|---|--|
| xxx xxx xxx | XXX XXX XXX | |

Bidders who have already submitted bids are hereby informed that they are allowed to modify or withdraw their bids, if necessary, before the scheduled opening of bid envelopes.

For modifications in your original submitted bid, kindly submit new bidding documents (sealed and marked as "Modified Bid") and have these received at the Office of the Bids and Awards Secretariat. Bid modifications received after the deadline shall not be considered and shall be returned to the bidder unopened.

ATTY. JOSEPHINE C. LATI-BAGAOISAN Chairperson



BID BULLETIN NO. 02

Date: 1

19 September 2022

ITB No.: bac-22-0907A

Project Name: Supply and Delivery of 2D Echo Machine – PCCH

ABC: Php9,630,679.00

To all prospective bidders:

This Bid Bulletin is issued to clarify, supplement, modify and/or revise the particular sections in the Bid and Contract Documents as stipulated in the Bidding Documents issued on 7 September 2022. The Bidders shall take note of the following items carefully and consider them in the preparation of their bid proposals, as they shall form part of the CONTRACT DOCUMENTS.

| Item | | revious Specification/ cation/Request to Consider | Amendment/Response to Clarification | | |
|------|---|---|---|---|--|
| 1 | PREVIOUS SECTION VII. TECHNICAL SPECIFICATIONS | | AMENDMENT TO SECTION VII. TECHNICAL SPECIFICATIONS | | |
| | | | | | |
| | ITEM | SPECIFICATION | ITEM | SPECIFICATION | |
| | 1 | 2D Echo Machine | 1 | 2D Echo Machine | |
| | | XXX XXX XXX | | XXX XXX XXX | |
| | | User Interface Operator Keyboard (minimum specifications or better) | | User Interface Operator Keyboard (minimum specifications or better) | |
| | | *Drawer type A/N keyboard (option) | | *Drawer type A/N keyboard (option) or Physical keyboard support for | |

| * Physical keyboard support for international characters in 7 languages (option) | international characters in 7 languages (option) |
|--|---|
| XXX XXX XXX | XXX XXX XXX |
| LCD Monitor (minimum specifications or better) | LCD Monitor (minimum specifications or better) |
| *256 shades of gray and 16.7 million simultaneous colors available | *256 shades of gray and 16.7 million simultaneous colors available or its equivalent |
| XXX XXX XXX | XXX XXX XXX |
| Operating Modes | Operating Modes |
| * 2D tissue | *2D tissue or its equivalent |
| * 2D angio flow | *2D angio flow or its equivalent |
| * Curved anatomical M- mode | * Curve <mark>d or</mark> anatomical M- mode |
| * Tissue synchronization imaging | * Tissue synchronization imaging or its equivalent |
| * AFI Automated Function Imaging | * AFI Automated Function Imaging or its equivalent |
| * Coded phase inversion | * Coded phase inversion or its equivalent |
| XXX XXX XXX | xxx xxx xxx |
| Transducers/Probes: (minimum specifications or better) | Transducers/Probes: (minimum specifications or better) |
| * Phased Array Adult – Active-Matrix Single Crystal | * Phased Array Adult – Active-Matrix Single Crystal or Single Crystal |
| XXX XXX XXX | XXX XXX XXX |
| Linear Array | Linear Array |